

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和7年9月22日

国立大学法人新潟大学長 牛木 辰男

◎調達機関番号 415 ◎所在地番号 15

○第3号

1 調達内容

- (1) 品目分類番号 22、31
- (2) 導入計画物品及び数量
生化学自動分析装置 一式
- (3) 調達方法 購入
- (4) 導入予定時期
令和8年度第2・四半期以降
- (5) 調達に必要とされる基本的な要求要件

1. システム構成

- A 生化学自動分析装置
- B 検体搬送システム

2. システム性能

(1) 生化学自動分析装置

- A 自動立ち上げ、自動後処理及び自動再検機能等を備えていること。
- B 生化学検査、電解質測定、尿分析を同時に行えること。
- C 検体搬送システムと接続可能であること。

(2) 検体搬送システム

- A 臨床化学検査等の検体について、開栓、分注等を行い、生化学自動分析装置等に自動的に検体の供給が行えること。
- B 検体前処理機能モジュールと各分析装置間を接続すること。

3. その他

- A 操作性及び安全性に優れていること。
- B 将来の機能拡張に対応できること。

2 資料及びコメントの提供方法 上記1(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

- (1) 資料等の提供期限 令和7年11月7日17時00分（郵送の場合は必着のこと。）

- (2) 提供先 〒951-8520 新潟市中央区旭町通1番町754番地 新潟大学医歯学総合病院管理運営課契約係 及川 俊哉 電話 025-227-0970
- 3 説明書の交付 本公表に基づき応募する供給者に対して導入説明書を交付する。
- (1) 交付期間 令和7年9月22日から令和7年11月7日まで。
- (2) 交付場所 上記2(2)に同じ。
- 4 説明会の開催 本公表に基づく導入説明会を開催する。
- (1) 開催日時 令和7年9月29日 15時00分
- (2) 開催場所 新潟大学医歯学総合病院西診療棟3階 第2会議室
- 5 その他 この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
- 6 Summary
- (1) Classification of the products to be procured : 22,31
- (2) Nature and quantity of the products to be purchased : Automatic biochemistry analyzer 1 Set
- (3) Type of the procurement : Purchase
- (4) Basic requirements of the procurement :
1. This system is constructed from the following modules for the purpose of :
- A Automatic biochemistry analyzer
- B Sample transportation system
2. Performance of the system:
- (1) Automatic biochemistry analyzer
- A The system shall be equipped with automatic start-up, automatic post-processing, and automatic re-examination functions.
- B The system shall be capable of performing biochemical testing, electrolyte measurement, and urinalysis simultaneously.
- C The system shall be connectable to the sample transportation system.
- (2) Sample transportation system
- A The system shall be capable of uncapping, dispensing, and automatically supplying specimens for clinical chemistry tests to the automatic biochemistry analyzer and other analytical instruments.
- B The system shall be capable of connecting the specimen preprocessing module with each analytical instrument included in the system configuration.
3. Other requirements :
- A The system shall be superior in terms of operability and safety.
- B The system shall be capable of accommodating future functional expansion.
- (5) Time limit for the submission of the requested material : 17:00 7 November,

2025

(6) Contact point for the notice : Toshiya Oikawa, Procurement Section, Niigata University Medical and Dental Hospital, 754 Asahimachidori 1-bancho Chuo-ku Niigata-shi 951-8520 Japan, TEL 025-227-0970